



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

SW

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/677,195	10/02/2003	David Croop		9624

7590

07/26/2004

R. Christian Macke
40 East 10th Street
Newport, KY 41071

EXAMINER

MANAHAN, TODD E

ART UNIT PAPER NUMBER

3732

DATE MAILED: 07/26/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/677,195

Applicant(s)

CROOP ET AL.

Examiner

Todd E. Manahan

Art Unit

3732

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-15 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-10 and 12-15 is/are rejected.
- 7) ☒ Claim(s) 11 is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 02 October 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. ____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date 01/12/04.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. ____.
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: ____.

Art Unit: 3732

DETAILED ACTION

Claim Rejections - 35 USC § 102

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-10, 12-15 are rejected under 35 U.S.C. 102(b) as being anticipated by Detsch (United States Patent No. 4,629,425).

Detsch discloses a dental mirror comprising an elongated handle; a reflective surface 104 affixed to a first end of the handle; means for communicating air through the handle to at least one air orifice 70 directing air at the reflective surface; means for communicating water through the handle to at least one water orifice 74 directing water at the reflective surface; an air valve 40 positioned on the handle for controlling air flow to the air orifice; and a water valve 50 positioned on the handle for controlling water flow to the water orifice. The reflective surface is attached to the handle by means of an externally threaded screw and an internally threaded sleeve.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-4 and 13 are rejected under 35 U.S.C. 103(a) as being unpatentable over Vellender (United States Patent No. 3,986,266 cited by applicant) in view of Detsch.

Vellender discloses the invention essentially as claimed except for the water valve means and the air valve means on the handle. Detsch discloses that it is known to provide such a device

Art Unit: 3732

with air and water valves on the handle in order that they may be manually depressed by one finger or thumb. It would have been obvious to one skilled in the art to provide the device of Vellender with air and water valves disposed on the handle in view of Detsch in order provide control of water and air at ones fingertips.

Claims 5-10 are rejected under 35 U.S.C. 103(a) as being unpatentable over Vellender in view of Detsch as applied to claims 1-4 above, and further in view of Whitman et al. (United States Patent No. 3,048, 924).

Whitman et al disclose a dental mirror device wherein the mirror is removably attached to the handle by an exteriorly threaded screw, interiorly threaded sleeve and lock nut. It would have been obvious to one skilled in the art to removably attach the mirror to the handle in view of Whitman et al. in order to permit the mirror to be ready replaced if damaged or broken.

Allowable Subject Matter

Claim 11 is objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Todd E. Manahan whose telephone number is 703 308-2695.

The examiner can normally be reached on Mon-Fri.


Art Unit: 3732

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Shaver can be reached on 703 308-2582. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Todd E. Manahan
Primary Examiner
Art Unit 3732

T. E. Manahan
22 July 2004

A handwritten signature in black ink, appearing to read 'T. E. Manahan', is written over the printed name and title. The signature is fluid and cursive, with a long horizontal stroke extending to the right.